

**From:** [Sharon Long](#)  
**To:** [Gallagher, Robert](#)  
**Cc:** [Sonya Brunty](#)  
**Subject:** RE: Request for Additional Information  
**Date:** Wednesday, May 01, 2013 5:03:14 PM  
**Attachments:** [image001.png](#)  
[Renewal Responce May 2013.pdf](#)

---

Hi Mr. Gallagher,

Attached is the signed NRC 313 form and a letter verifying access to the building. As noted below there is nothing above or below the department with the exception of the gamma camera room. No radioactive materials are store in this area. Additionally surveys of the floor area of the camera room showed exposures to be < 2mr/hr.

As a note, the Building where the cardiology office is located now has a new name. It was previously named Robert C. Byrd Clinical Center. It is now named Erma Ora Byrd Clinical Center. There is no physical change in the address of use. Please update the license to reflect this change.

Please let me know if there is anything else that you need.

*Sharon L. Long*  
*National Physics Consultants*  
*1296 Russell Drive*  
*Bolivar, Oh 44612*  
*Phone – 888-456-5255*  
*Fax- 866-872-9302*  
Sharon,

Attached are the documents that Sonya needed to send to you.

And about the question as to what is below the lab:

There is nothing above our floor  
Nothing is below the prep room, treadmill room and hot lab  
There are exam rooms under the camera room

*Nancy Floyd*  
Program Coordinator  
Adult Cardiology Training Fellowship  
Interventional Cardiology Fellowship

1249 15<sup>th</sup> Street, Suite 4000  
Huntington, WV 25701  
304-691-8534 – phone  
304-691-8530 - fax

---

**From:** Gallagher, Robert [mailto:Robert.Gallagher@nrc.gov]  
**Sent:** Wednesday, April 24, 2013 1:26 PM  
**To:** slong@nationalphysics.com  
**Subject:** Request for Additional Information

**PLEASE CONFIRM RECEIPT OF THIS REQUEST FOR ADDITIONAL INFORMATION BY RETURN EMAIL**

License No. 47-25620-01  
Docket No. 03036212  
Control No. 579806

Dear Ms. Long;

This is in reference to your application [NRC Form 313] dated January 15, 2013 requesting to renew Nuclear Regulatory Commission License No. 47-25620-01. In order to continue our review, we need the following additional information:

1. Your application should have been signed by a management representative. Please submit NRC Form 313 signed by a management representative indicating that management has reviewed the application and concurs in the statements and representations contained therein. Note also that a management representative should sign all future correspondence that requests a change in your license.
2. Your application appears to indicate that your location of use may be controlled by an entity other than yourself. If so, please provide documentation of a clear contractual agreement concerning access to your location of use for the purpose of decontamination or removal of licensed material from the location of use in the event of disharmony between you and the owner entity. This documentation should consist of signed certification from both parties.
3. Please provide a description of the areas above and below your use and storage areas and show that adequate steps have been taken to assure that radiation levels in unrestricted areas will not result in doses to individual members of the public in excess of those specified in 10 CFR 20.1301.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 A.M. to 5:30 P.M., Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 579806. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5182.

In order to continue prompt review of your application, we request that you submit your response to this

letter within 30 calendar days from the date of this letter.

Sincerely,

Robert L. Gallagher  
Health Physicist  
U.S. NRC, Region I  
2100 Renaissance Blvd.  
King of Prussia, PA 19406  
(610) 337-5182 office  
(610) 337-5269 fax

NRC Seal

